



Infection Prevention with Confidence!



In the war against microbes, Accel® Disinfectants are the first line of defense for Infection Prevention. Accel® provides comfort and confidence that commonly touched surfaces and instruments are rendered safe to use.

Accel CS 20 is designed for use in Healthcare and Personal Services when sterilization in a rapid contact time is required. Accel CS 20 achieves chemical sterilization and high level disinfection in 20 minutes.

EFFICACY OF ACCEL CS 20

SPORICIDAL (20 min at 20°C): *Bacillus subtilis*, *Clostridium sporogenes*

TUBERCULOCIDAL (20 min at 20°C): *Mycobacterium terrae*

FUNGICIDAL (20 min at 20°C): *Trychophyton mentagrophytes*

DIRECTIONS FOR USE

(INSTRUMENTS REQUIRE CLEANING PRIOR TO DISINFECTION)

CRITICAL ITEMS: Chemosterilization for instruments and devices that come in contact with sterile tissue, the blood stream & recirculating body fluids – such as surgical instruments and foot care instruments. Place pre-cleaned, rinsed and dried instruments into a container of undiluted Accel CS 20. Allow the instruments to soak for at least 20 minutes. Remove and rinse with sterile water. (See Notice To User)

SEMI-CRITICAL ITEMS: High-Level Disinfection for instruments and devices that come in contact with non-intact skin or mucous membranes that are not normally considered sterile such as respiratory equipment, surgical mirrors etc. Place pre-cleaned, rinsed and dried instruments into a container of undiluted Accel CS 20. Allow the instruments to soak for at least 20 minutes. Remove and rinse with sterile or potable water. (See Notice To User)

14 DAY RE-USE CLAIM: This solution may be used and re-used for up to 14 days. Solution in its original packaging is good until date of expiry. Test kits (AHP7, K5500C) for confirming the active ingredient and the stability of Accel CS 20 are available.

NOTICE TO USER: Please ensure the instrument, equipment or device material to be disinfected is compatible with Accel CS 20. Compatible materials include: acrylic, silicone rubber, low and high density plastics, rigid stainless steel and mild steel instruments. Continuous exposure to copper and brass should be avoided. Accel CS 20 (visit the Compatibility section on www.virox.com for a full list of compatible materials) is not recommended for use on carbon steel and anodized aluminum and is not intended for use on flexible endoscopes or intracavity ultrasound probes. If you are uncertain as to the material composition of your item, confirm with the manufacturer before proceeding.

Accel CS 20

Product Number	ACCCS20-1, ACCCS20-4
DIN #	02281473
Concentration of Active	7% Hydrogen Peroxide
Container Size	1L, 4L
Dilution	Ready to Use
Shelf Life	12 Months



Engineering Revolutionary Disinfectants for the War Against Microbes

1 (800) 387-7578
www.virox.com