



## Accel WASH Directions for Use

### Protective Barriers:

1. Disposable gloves. Gloves should be changed as required, i.e., when torn, when hands become wet inside the glove or when moving between patient rooms.
2. Household gloves can be worn, but they must be discarded when the cleaning is complete.
3. Protective Eye wear (goggles, face shield or mask with eye protection) should be worn when mixing concentrates.
4. Masks are not normally required if good ventilation is maintained and exposure guidelines are not exceeded.
5. Gowns should be used when working with concentrates.

### Dilutions:

Determine method for the dispensing of Accel WASH Concentrate at a ratio of 1:64 for Ultrasonic Cleaning, 1:128 for Manual, Cart or Automatic Machine Cleaning, 1:256 for light manual cleaning. Accel WASH carries a 24 month shelf life from the manufacturing date as listed on the original container.

<i>Ratio</i>	<i>Per US Gallon</i>	<i>Per 4 L Water</i>
1:256	½ oz.	16 ml
1:128	1 oz.	32 ml
1:64	2 oz.	64 ml

\*Once diluted, Accel WASH should be discarded at end of day or when solution becomes visibly soiled.

### Recommended Procedures for Initial Manual Cleaning:

1. Dilute Accel WASH at 1:128 or 1:256 at 20°C into a container.
2. Place instruments in solution for 3 to 5 minutes using agitation when required.
3. Once contact time has elapsed, remove instruments from container.
4. Rinse instruments with water.
5. Inspect instruments thoroughly for cleanliness.
6. Once found to be clean, dry and proceed with appropriate disinfection step.
7. Change solution daily or as needed.
8. Accel WASH is not intended for use as a holding solution for non-critical, semi-critical or critical items.

### Recommended Procedures for Ultrasonic Cleaning:

1. Dilute Accel WASH at 1:64 at 20°C into an ultrasonic machine holding tank.
2. The use of hot water is not recommended.
3. Place the instruments into the ultrasonic machine and run for 3 to 5 minutes.
4. Remove the instruments from the machine.
5. Rinse instruments with water.
6. Inspect instruments thoroughly for cleanliness.



## ***Accel WASH*** **Directions for Use**

7. Once found to be clean, dry and proceed with appropriate disinfection step.
8. Change solution daily or as needed.

### **Recommended Procedures for Automatic Machine Reprocessing:**

1. Dosing may be manually or automatically fed.
2. For use with heated water up to 60°C (140°F).
3. Dilute Accel WASH at 1:128.
4. Place the instruments into the machine for a 5 minute cycle time.
5. Remove the instruments from the machine.
6. Rinse instruments with water.
7. Inspect instruments thoroughly for cleanliness.
8. Once found to be clean, dry and proceed with appropriate disinfection step.

### **Recommended Procedures for Cart Washing:**

1. Dosing may be manually or automatically fed.
2. For use with heated water up to 60°C (140°F).
3. Dilute Accel WASH at 1:128.
4. Place carts through the machine for a 5 minute wash cycle.
5. Remove the cart from the machine.
6. Rinse cart with water.
7. Allow to air dry.