



Accel CS20 Directions for Use

Protective Barriers:

1. Disposable gloves. Gloves should be changed as required, i.e., when torn, when hands become wet inside the glove.
2. Household gloves can be worn, but they must be discarded when the cleaning is complete.
3. Protective Eye wear (goggles, face shield or mask with eye protection) should be worn when decanting.

Dilutions:

Product is Ready to Use and does not require dilution.

Shelf Life: 12 months from date of manufacture. Expiry date is listed on each product label. Accel CS20 carries a 14 day re-use period.

Recommended Procedures for Cleaning and Disinfecting of Equipment:

1. Cleaning should take place between each device/equipment usage.
2. Items are to be disassembled (as appropriate) and thoroughly cleaned.
3. Items may be contaminated with dried or wet sputum and/or blood and should be cleaned using a detergent such as Accel WASH (Diluted 1:128 or 1:256), rinsed, and dried prior to Accel CS20 for High Level Disinfection/Chemical Sterilization.
4. Disinfection consists of immersing the device/equipment in the Accel CS20 solution for 20-minutes.
5. Once the 20-minute contact time has passed, the device/equipment should be removed from the Accel CS20 solution using clean, disinfected tongs or forceps.
6. The disinfected device/equipment should be thoroughly rinsed using sterile water when practical, otherwise potable water is acceptable for semi-critical devices.
7. Following removal from Accel CS20 solution, thoroughly rinse device/equipment by immersing it completely in a large volume of water keeping the device/equipment totally immersed for a minimum of 1-minute repeating this step for 3 consecutive times following CSA Guidelines (pg. 43 - .
8. Manually flush all lumens.
9. Remove device/equipment and discard the rinse water. Always use fresh water for each rinse. Do not reuse the water for rinsing or any other purposes.
10. Following the rinsing step the device/equipment should be dried and stored in a suitable container (i.e. Clear polyethylene bag) for future use.



Accel CS20 **Directions for Use**

Instructions for Confirmatory Testing of Accel CS20 (7% AHP):

The Accelerated Hydrogen Peroxide Test Strip 7% (Part No. AHP7) should be used for confirmatory testing. These strips are easy to use dip-and-read reagents strips for a pass or fail determination of the hydrogen peroxide concentration in the Accel CS20 solution.

1. Remove a test strip and immediately close the container.
2. Dip the test strip into the Accel CS20 solution to be tested for 1-second ensuring that the reaction zone is completely wetted.
3. Remove the test strip and shake off excess liquid.
4. Wait for 35-seconds then compare the reaction zone with the colour scale.

NOTE: The purpose of confirmatory testing is not to extend the reuse life beyond the 14-day claim. Should the test strip show that the Accel CS20 solution still meets the targeted level of hydrogen peroxide the product **MUST** still be disposed.